

Comparison of measured Serum Total Carbon Dioxide with calculated Bicarbonate calculated from Arterial Blood Gas Analysis

¹Thuthi Mohan, ²B Vinodh Kumar

ABSTRACT

Introduction: Measured total carbon dioxide (TCO₂) from venous sample and calculated bicarbonate from arterial blood gas (ABG) have shown good agreement in some studies, while conflicting results have been obtained in few other studies. The objective of this study is to compare and assess the degree of agreement between the measured TCO₂ and calculated bicarbonate and also whether they can be used interchangeably in our laboratory.

Materials and methods: We prospectively analyzed 89 ABG samples requested for calculated bicarbonate and then measured TCO₂ from venous blood samples drawn simultaneously from the same participants between November 2016 and April 2017.

Results: Measured TCO₂ results ranged from 5.7 to 39.9 mmol/L (mean 23.45 mmol/L), while calculated bicarbonate ranged from 9 to 40 mmol/L (mean 24.36 mmol/L). The values of TCO₂ and bicarbonate correlated well ($r = 0.95$, $p < 0.001$), with the correlation given by the equation, $y = 0.884x + 3.605$. The bias obtained was -0.9 mmol/L and the standard deviation (SD) was 1.62 mmol/L. The limits of agreement (LOA) were -4.1 to 2.3 mmol/L, with a span of 6.4 mmol/L. Out of the 89 values, 85 (95.05%) were within the LOA.

Conclusion: In majority of the cases, the calculated bicarbonate concentration from ABG showed a good correlation to the measured venous TCO₂ concentration. Despite this excellent correlation, TCO₂ did not show good agreement with calculated bicarbonate when Story and Poustie's criteria were applied, especially in cases of bicarbonate less than 20 mmol/L. Hence, clinicians should be aware of this discrepancy and be cautious when using measured TCO₂ and calculated bicarbonate interchangeably in the assessment and management of acid–base disorders, especially in patients with metabolic acidosis.

Keywords: Arterial blood gas, Bicarbonate, Henderson–Hasselbalch equation, Metabolic acidosis, Total carbon dioxide.

How to cite this article: Mohan T, Kumar BV. Comparison of measured Serum Total Carbon Dioxide with calculated

Bicarbonate calculated from Arterial Blood Gas Analysis. Indian J Med Biochem 2017;21(2):76-80.

Source of support: Nil

Conflict of interest: None

INTRODUCTION

Measurement of TCO₂ is useful in determining the acid–base status of a patient. The TCO₂ in serum or plasma exists in three major chemical forms: 95% as bicarbonate anion, 3% as dissolved CO₂, and 2% as carbamino derivatives of plasma protein.¹ Other quantitatively minor forms are carbonic acid and carbonate ions. The use of the term “bicarbonate” to describe the measured parameter is not correct since the measured TCO₂ includes dissolved CO₂, carbamino compounds, and carbonic acid and not strictly bicarbonate. Some laboratories use the term “TCO₂” to describe the measured bicarbonate. Two commonly employed methods to quantitate TCO₂ content in serum or plasma involve acidification of the sample to convert all carbon dioxide forms to CO₂ gas and measure the amount of gas formed using PCO₂ electrode. The second method utilizes the enzymatic reaction of bicarbonate with phosphoenolpyruvate carboxylase coupled to a spectrophotometric indicator reaction. Most assays that specifically measure bicarbonate have reagents at an alkaline pH. This converts other forms of CO₂ to bicarbonate, so these assays are also measuring total CO₂. Bicarbonate can also be calculated from blood–gas analysis using the Henderson–Hasselbalch equation. The equation used to calculate bicarbonate is $[\text{HCO}_3^-] = \alpha \text{PCO}_2 (\text{antilog} [\text{pH} - \text{pKa}])$. This equation is based on the assumption that the dissociation constant (pKa) and solubility coefficient (α) are 6.1 and 0.03 respectively. However, pKa is affected by changes in pH, ionic strength, and temperature while the values of solubility coefficient (α) vary with the composition of the solution, such as the presence of increased salts proteins or lipids.² Therefore, the calculated bicarbonate values may have significant error under certain circumstances irrespective of the sample type, making the reliability questionable. Previously done studies to assess the agreement between calculated bicarbonate and measured TCO₂ have shown good agreement in some studies³⁻⁵ while revealing

¹Assistant Professor, ²Senior Resident

^{1,2}Department of Biochemistry, ESIC Medical College & Postgraduate Institute of Medical Sciences and Research Chennai, Tamil Nadu, India

Corresponding Author: B Vinodh Kumar, Senior Resident Department of Biochemistry, ESIC Medical College & Postgraduate Institute of Medical Sciences and Research Chennai, Tamil Nadu, India, Phone: +919840652990, e-mail: drvinodh87@gmail.com

conflicting results in others.⁶⁻⁸ The objective of this study is to compare and assess the degree of agreement between the measured TCO₂ and calculated bicarbonate and also whether they can be used interchangeably.

MATERIALS AND METHODS

We prospectively analyzed 89 ABG samples requested for calculated bicarbonate and measured TCO₂ from venous blood samples drawn simultaneously from the same participants between November 2016–April 2017. Out of 89 participants, 54 were males within the age range (9–68 years) and 35 were females within the age range (11–66 years). Measured TCO₂ results ranged from 5.7–39.9 mmol/L (mean 23.45 mmol/L), while calculated bicarbonate ranged from 9–40 mmol/L (mean 24.36 mmol/L).

The TCO₂ was analyzed in venous sample by the Vitros ECO₂ Slide method in Vitros 4600 Chemistry System analyzer.⁹ The bicarbonate reagent slide utilizes the enzymatic method whereby the high pH in the spreading layer ensures that essentially all CO₂ in the sample is in the bicarbonate form. The bicarbonate then diffuses to the gel layer and is used to carboxylate phosphoenolpyruvate in the presence of phosphoenolpyruvate carboxylase to form oxaloacetate and inorganic phosphate. The final reaction involves the malate dehydrogenase-catalyzed oxidation of nicotinamide adenine dinucleotide (NADH) and reduction of oxaloacetate to produce NAD⁺ and malate. The slide is incubated at 37°C. The concentration of CO₂ in the sample is determined by measuring the absorbance of the unreacted NADH by reflectance spectrophotometry.

Bicarbonate was calculated by the Roche Cobas b 221 Blood–Gas Analyzer using the Henderson–Hasselbalch equation. The equation used to calculate bicarbonate is as follows $[\text{HCO}_3^-] = \alpha \text{PCO}_2 (\text{antilog} [\text{pH} - \text{pKa}])$. This equation is based on the assumption that pKa is the dissociation constant for carbonic acid (equal to 6.1 for blood at 37°C) and α is the solubility coefficient for CO₂ gas (equal to 0.03 for plasma at 37°C).

Statistical Analysis

The statistical analysis was performed using Microsoft Excel 2007.

(a) Paired t-test was used to compare all participants included in the study between the calculated bicarbonate from ABG and measured TCO₂. The paired t-test was used since all parameters used in the study are numerical, all participants came from same population, and all had paired results.

(b) The comparison plot/least square linear regression analysis was used for comparison of calculated bicarbonate and measured TCO₂. The values obtained

by the measured TCO₂ are plotted on the y-axis and the calculated bicarbonate from ABG is plotted on the x-axis. A line of “best fit” is drawn in such a way that there are minimum squares of distance of all the data points from the line chosen. This line is called “least squares line”. This graph/line has the following equation: $y = a + bx$; where “y” is test method, where “a” is called y intercept (the point/value at which the line cuts the y axis or y value).

(c) The Bland–Altman plot/difference plot was used for comparison of calculated bicarbonate and measured TCO₂. This was analyzed with the averages of measured TCO₂ and calculated bicarbonate on the x-axis and the differences between the two in mmol/L, plotted on the y-axis.

RESULTS


Table 1 shows the comparison of all participants included in the study by Paired t-test, between the calculated bicarbonate from ABG and measured TCO₂. A statistically significant difference ($p < 0.05$) was observed between calculated bicarbonate from ABG and measured TCO₂ when all participants included in the study independent of age, sex, and race were compared.

Graph 1 shows the comparison of calculated bicarbonate and measured TCO₂ by comparison plot/least


Table 1: Comparison of all participants included in the study by paired t-test between the calculated bicarbonate and measured TCO₂

	Calculated bicarbonate from ABG (arterial blood)	Measured TCO ₂ —venous blood	p-value
Total no of participants	89	89	<0.001
Mean	23.45	24.36	
SD	7.01	6.36	
SE	2.48	2.58	

SD: Standard deviation; SE: Standard error



Graph 1: Comparison of calculated bicarbonate and measured TCO₂ by comparison plot/least square linear regression analysis



Graph 2: The Bland–Altman plot/difference plot analyzed with the averages of measured TCO₂ and calculated bicarbonate on the x-axis and the differences between the two in mmol/L are plotted on the y-axis

square linear regression analysis. The values obtained by the measured TCO₂ are plotted on the y-axis and the calculated bicarbonate from ABG is plotted on the x-axis. Measured TCO₂ results ranged from 5.7–39.9 mmol/L (mean 23.45 mmol/L), while calculated bicarbonate ranged from 9–40 mmol/L (mean 24.36 mmol/L). The values of TCO₂ and bicarbonate correlated well ($r = 0.95$,

$p < 0.001$), with the correlation given by the equation, $y = 0.884x + 3.605$.


Graph 2 shows the Bland–Altman plot/difference plot. This was analyzed with the averages of measured TCO₂ and calculated bicarbonate on the x-axis and the differences between the two in mmol/L, plotted on the y-axis. The upper and lower LOA were calculated from bias ± 1.96 SD. The span was calculated from the high limit and low limit values. The bias obtained was -0.9 mmol/L and the SD was 1.62 mmol/L. The LOA were -4.1 to 2.3 mmol/L, with a span of 6.4 mmol/L. Out of the 89 values, 85 (95.05%) were within the LOA.

Table 2 shows the comparison of participants divided into 3 subgroups based on the average value of bicarbonate [<20 mmol/L (group I), $20\text{--}26$ mmol/L (group II) and >26 mmol/L (group III)] included in the study by paired t-test between the calculated bicarbonate and measured TCO₂. A statistically significant difference ($p < 0.001$) was observed between calculated bicarbonate and measured TCO₂ in the groups I and II, but a statistically insignificant p-value ($p\text{-value} > 0.05$) was observed in the group III.


Graphs 3 to 5 show the Bland–Altman plot analysis with the average value of measured TCO₂ and calculated bicarbonate [< 20 mmol/L (group I), $20\text{--}26$ mmol/L (group II), and $26\text{--}40$ mmol/L (group III)] on the x-axis

Table 2: Comparison of participants divided into three subgroups based on the average value of bicarbonate <20 mmol/L (group I), $20\text{--}26$ mmol/L (group II), and >26 mmol/L (group III) included in the study by paired t-test between the calculated bicarbonate and measured TCO₂


Average bicarbonate (mmol/L)	No. of participants	Mean of calculated bicarbonate	Mean of measured TCO ₂	Average mean	p-value
<20	24	15.15	17.15	16.15	<0.001
$20\text{--}26$	33	22.34	22.98	22.66	0.01
>26	32	30.83	31.18	31.01	0.16



Graph 3: The Bland–Altman plot analysis with the averages of measured TCO₂ and calculated bicarbonate (<20 mmol/L) on the x-axis and the differences between the two in mmol/L are plotted on the y-axis



Graph 4: The Bland–Altman plot analysis with the averages of measured TCO₂ and calculated bicarbonate ($20\text{--}26$ mmol/L) on the x-axis and the differences between the two in mmol/L are plotted on the y-axis



Graph 5: The Bland–Altman plot analysis with the averages of measured TCO₂ and calculated bicarbonate (>26 mmol/L) on the x-axis and the differences between the two in mmol/L are plotted on the y-axis

and the differences between the two in mmol/L plotted on the y-axis respectively.

DISCUSSION

The agreement between calculated bicarbonate from venous sample and measured TCO₂ from ABG have shown good agreement in some studies³⁻⁵ while revealing conflicting results in others.⁶⁻⁸ The objective is to find out whether the values obtained by both the methods can be used interchangeably in our laboratory.

Theoretically measured TCO₂ is slightly higher than calculated bicarbonate. This is because the measured TCO₂ includes an additional nonbicarbonate CO₂ component which amounts to around 1.2 mmol/L. The other reason is that the serum or plasma sample used to measure TCO₂ is derived from venous blood, which in itself has a slightly higher bicarbonate concentration than that of the arterial blood. These theoretical reasons in themselves amount to approximately 2.0 mmol/L difference between calculated and measured bicarbonate.¹⁰⁻¹²

The first step in assessing agreement between measured and calculated bicarbonate by correlation coefficient showed a good correlation with $r = 0.950$ with $p < 0.05$. The correlation coefficient to assess the agreement depends not only on the correlation coefficient, but also on the range of values, so a wide range of values will show a good correlation even if the values are in poor agreement.

By Bland–Altman analysis, the bias and SD obtained were -0.9 and 1.62 mmol/L respectively. The LOA were -4.1 to 2.3 mmol/L, with a span of 6.4 mmol/L. Out of the 89 values, 85 (95.05%) were within the LOA. Then Story and Poustie's criteria were used to assess the degree of agreement between calculated bicarbonate and measured TCO₂.¹³ The criteria proposed for agreement are:

(i) The bias or difference between calculated TCO₂ and measured bicarbonate should be less than ± 1 mmol/L, (ii) the LOA between the methods should range within a bias ± 2 mmol/L or within a total span of 4 mmol/L, to be clinically insignificant. The results of our study showed a bias of -0.9 mmol/L, which fulfilled the first criteria, but the LOA were -4.1 to 2.3 mmol/L, with a span of 6.4 mmol/L, which did not fulfill the second criteria. Despite good statistical correlation obtained by paired t-test and linear regression analysis, the application of criteria proposed by Story and Poustie for agreement between TCO₂ and HCO₃ is not fully satisfied.

The participants were subgrouped into three groups based on the average value of measured TCO₂ and calculated bicarbonate as follows, average bicarbonate < 20 mmol/L (group I), 20 – 26 mmol/L (group II) and 26 – 40 mmol/L (group III) independent of age and sex. On application of the criteria of Story and Poustie, the first subgroup did not fulfill both the criteria but the 2nd and 3rd subgroups fulfilled the first criteria, but not the second criteria. Hence, the poor agreement between the two methods was observed at lower bicarbonate concentration, especially at bicarbonate concentration at < 20 mmol/L.

The discrepancy in values between measured TCO₂ and calculated bicarbonate can be explained because of usage of a constant pKa of 6.1 in the Henderson–Hasselbalch equation, i.e., used to calculate the bicarbonate. The study of Flear et al¹⁴ showed that pKa values varied considerably by more than 0.06 in healthy volunteers and in very ill patients and hence, an error of approximately $\pm 60\%$ was noted in calculated bicarbonate based on constant pKa of 6.1. Another study by O'Leary and Langton found a significant decrease of the pKa in patients considered to have metabolic acidosis when compared with patients with bicarbonate concentration within the reference interval.⁶ The findings in our study also noted that majority of values falling outside the LOA occurred at bicarbonate concentration < 20 mmol/L, which seem to be in agreement with O'Leary and Langton.

The other cause of discrepancy in results could be due to preanalytical factors, such as sample collection and handling like inadequate mixing, acidification, non-maintenance of anaerobic condition, and dilution of ABG samples with excess heparin that can decrease the pH and pCO₂, and, hence, will affect the bicarbonate values.^{6,7} The clinicians would prefer instantly the calculated bicarbonate results from the ABG instrument placed in the ICU for the management of acid–base disorders, rather than sending for measured TCO₂ to the laboratory. But the findings of poor agreement would have an impact on the utility of calculated bicarbonate by the clinicians, especially, in

patients with metabolic acidosis. Parameters which are calculated using bicarbonate concentration like base excess and anion gap may also be affected by this discrepancy.

CONCLUSION

In the majority of cases, the calculated bicarbonate concentration from ABG showed a good correlation to the measured venous TCO₂ concentration. Despite the statistical correlation, TCO₂ did not show good agreement with calculated bicarbonate when Story and Poustie criteria were applied, especially in cases of bicarbonate less than 20 mmol/L. Hence, it is necessary to be cautious when using measured TCO₂ and calculated bicarbonate interchangeably in the assessment and management of acid-base disorders, especially in patients with metabolic acidosis.

REFERENCES

1. Davenport HW. The ABC of acid-base chemistry. 5th ed. Chicago: University of Chicago Press; 1969. p. 34.
2. Scott MG, Klutts JS. Electrolytes and blood gases. In: Burtis CA, Ashwood ER, Bruns DE, editors. Tietz textbook of clinical chemistry and molecular diagnostics. 4th ed. St. Louis (MO): Saunders; 2006. p. 983-1018.
3. Kumar V, Karon BS. Comparison of measured and calculated bicarbonate values. Clin Chem 2008 Sep;54(9):1586-1587.
4. Chittamma A, Vanavanan S. Comparative study of calculated and measured total carbon dioxide. Clin Chem Lab Med 2008;46(1):15-17.
5. Ungerer JP, Ungerer MJ, Vermask WJ. Discordance between measured and calculated total carbon dioxide. Clin Chem 1990 Dec;36(12):2093-2096.
6. Story DA, Poustie S, Bellomo R. Comparison of three methods to estimate plasma bicarbonate in critically ill patients: Henderson-Hasselbalch, enzymatic and strong-ion-gap. Anaesth Intensive Care 2001 Dec;29(6):585-590.
7. Lolekha PH, Boonlert W, Kost GJ, Vanavanan S, Lolekha S. Comparative study of values of calculated bicarbonate and measured total carbon dioxide content. Point Care 2003 Jun;2(2):135-143.
8. O'Leary TD, Langton SR. Calculated bicarbonate or total carbon dioxide? Clin Chem 1989 Aug;35(8):1697-1700.
9. Fezzotti A, Gambini AM, Coppa G, De Sio G. Total carbon dioxide measured by the Vitros enzymatic method. Clin Chem Lab Med 1998 Jan;36(1):43-46.
10. Kelly AM, McAlpine R, Kyle E. Agreement between bicarbonate measured on arterial and venous blood gases. Emerg Med Australas 2004 Oct-Dec;16(5-6):407-409.
11. Middleton P, Kelly AM, Brown J, Robertson M. Agreement between arterial and central venous values for pH, bicarbonate, base excess and lactate. Emerg Med J 2006 Aug;23(8):622-624.
12. Malatesha G, Singh NK, Bharija A, Rehani B, Goel A. Comparison of arterial and venous pH, bicarbonate, pCO₂ and pO₂ in initial emergency department assessment. Emerg Med J 2007 Aug;24(8):569-571.
13. Story DA, Poustie S. Agreement between two plasma bicarbonate assays in critically ill patients. Anaesth Intensive Care 2000 Aug;28(4):399-402.
14. Fleary CT, Roberts SW, Hayes S, Stoddart JC, Covington AK. pK₁' and bicarbonate concentration in plasma. Clin Chem 1987 Jan;33(1):13-20.